



N42RF ERROR SUMMARY

SENEX 2013 KMQY

02 July 2013



Flight ID: 20130702H1

<u>Sensor or system</u>	<u>Number or Name</u>
Static Pressure Probe	PSM.2
Dynamic Pressure Probe	PQM.2
Total Temperature Probe	TTM.1
Dewpoint Temp. Probe	TDM.2
Vertical Accelerometer	AccZfilterI-GPS.1
Altimeter	AltIGPS.1
INE Selection	1
Differential Attack Pressure Probe	PDALPHA.1
Differential Sideslip Pressure Probe	PDBETA.1
Dynamic Attack Pressure Probe	PQALPHA.1
Dynamic Sideslip Pressure Probe	PQBETA.1
Flight Directory	acdata/MET/2013/20130702H1

Local Met Data:	<u>Takeoff</u> (0056Z) 03July	<u>Landing</u> (0754Z) 03 July
Aircraft Static Pressure	998.7mb	998.0mb
Tower Pressure (corrected)	997.1mb	998.1mb

Notes:

All AOC instruments worked properly.

After the aircraft returned to MacDill AFB, a few instruments that were installed on N42RF were removed and calibrated in the AOC calibration laboratory. The final Quality Control data sets use the post project calibration coefficients for determining the output of the instruments.

The following parameters that have been updated from the original data set follow.

PDALPHA.2 - Radome Attack Differential Pressure Measured
PDBETA.2 - Radome Sideslip Differential Pressure Measured
PDALPHA.1 - Fuselage Attack Differential Pressure Measured
PQALPHA.1 - Fuselage Attack Dynamic Pressure Measured
PDBETA.1 - Fuselage Sideslip Differential Pressure Measured
PQBETA.1 - Fuselage Attack Dynamic Pressure Measured
PQM.3 - Measured Fuselage Dynamic Pressure, Two Sources --co-pilot
PQM.4 - Radome Dynamic Pressure Measured
TTM.1 - Measured Total Temperature Degree
TTM.2 - Measured Total Temperature Degree

Any derived value using the outputs from the measured parameters listed above have changed slightly from the values measured in real-time.

See SENEX_READ_ME.txt for file information.

Flight Director:
Phone #:

Ian Sears
(813) 828-3310 ext. 3039